


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1944, Rack A, Plate 1, Well D07, 170582
[D8701C6304CBA4CA] Oct 10, 2025 11:48:56 Pressure = 67 bar (67 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.48	1.0%	9436	
2	0.488	F 0.58	0.5%	5030	
3	0.677	F 0.81	9.3%	88099	Acetylated F peak
4	0.832	F 0.99	52.0%	493788	Consistent with F
5	0.988	F 1.18	1.9%	18160	
6	1.202	A 1.01	34.8%	330340	A peak - REVIEW
7	1.368	A 1.15	0.6%	5619	
Total Area:				950472	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range