


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2122, Rack A, Plate 1, Well A06, 489321
 [CC2352027E0BD937] Apr 13, 2026 14:14:30 Pressure = 59 bar (57 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.22	4.1%	47988	
2	0.602	F 0.31	5.2%	60872	
3	0.898	F 0.46	16.9%	199307	
4	1.135	F 0.59	5.0%	59325	
5	1.883	F 0.97	58.2%	685982	Consistent with F
6	2.518	F 1.30	2.7%	31951	
7	2.788	F 1.44	0.6%	7585	
8	3.155	A 0.84	0.2%	2261	
9	3.678	A 0.98	7.1%	83873	A peak
				Total Area: 1179144	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range