


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1911, Rack A, Plate 1, Well C04, 455713  
 [C9355138B1232CC0] Sep 17, 2025 12:56:43 Pressure = 60 bar (60 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.50	1.1%	13564	
2	0.495	F 0.59	1.3%	16009	
3	0.697	F 0.84	10.1%	120037	Acetylated F peak
4	0.842	F 1.01	73.9%	879372	Consistent with F
5	1.062	A 0.89	0.4%	4619	
6	1.203	A 1.01	13.1%	156001	A peak
Total Area:				1189602	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range