


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1911, Rack A, Plate 1, Well A05, 170127
[D9250038B3812462] Sep 17, 2025 12:10:14 Pressure = 61 bar (61 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.50	0.7%	6722	
2	0.497	F 0.60	0.8%	7197	
3	0.698	F 0.84	14.8%	133971	Acetylated F peak
4	0.833	F 1.00	70.1%	633061	Consistent with F
5	0.962	F 1.15	1.5%	13280	
6	1.195	A 1.00	12.1%	109184	A peak
Total Area:				903415	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range