


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2146, Rack A, Plate 1, Well E07, 496542
[8C29F0C5FF05BB1F] May 13, 2026 13:32:32 Pressure = 65 bar (65 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.518	F 0.61	1.7%	22607	
2	0.710	F 0.83	13.7%	178148	Acetylated F peak
3	0.852	F 1.00	65.1%	843947	Consistent with F
4	1.227	A 1.00	19.4%	250980	A peak
Total Area:				1295682	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range