


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2121, Rack A, Plate 1, Well H04, 172988
[C9734603017624CA] Apr 13, 2026 13:31:20 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.48	1.2%	12302	
2	0.512	F 0.59	1.2%	12162	
3	0.722	F 0.83	14.6%	153296	Acetylated F peak
4	0.883	F 1.01	65.8%	691695	Consistent with F
5	1.243	A 1.00	17.3%	182081	A peak

Total Area: 1051536

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range