


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2111, Rack A, Plate 1, Well E09, 172852
 [9966044554DEACE0] Mar 31, 2026 16:13:03 Pressure = 60 bar (58 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.21	6.8%	98341	
2	0.543	F 0.27	15.5%	225000	
3	0.950	F 0.48	23.8%	344517	
4	1.982	F 1.00	25.5%	368492	Consistent with F
5	2.340	F 1.18	0.3%	5050	
6	2.525	F 1.27	0.6%	9170	
7	2.750	F 1.38	2.5%	36681	
8	2.983	A 0.80	1.1%	16541	
9	3.275	A 0.88	0.8%	11191	
10	3.752	A 1.00	22.1%	319311	A peak - REVIEW
11	4.273	A 1.14	0.4%	5807	
12	4.477	A 1.20	0.3%	4857	
13	4.867	S 0.92	0.2%	2464	
				Total Area: 1447422	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range