


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2077, Rack A, Plate 1, Well D08, 489345  
[987723D2F779D997] Feb 18, 2026 11:15:22 Pressure = 56 bar (55 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.47	1.5%	14549	
2	0.512	F 0.58	0.9%	9346	
3	0.720	F 0.82	14.1%	140770	Acetylated F peak
4	0.867	F 0.99	49.7%	497193	Consistent with F
5	1.023	F 1.17	3.2%	32445	
6	1.257	A 1.00	30.6%	306495	A peak - REVIEW
Total Area:				1000798	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range