


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2140, Rack A, Plate 1, Well C12, 496417 FA
 [99764E2E64BC0EE2] May 06, 2026 12:40:30 Pressure = 64 bar (63 to 66)



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.433 | F 0.51 | 1.1% | 17848 | |
| 2 | 0.513 | F 0.60 | 0.4% | 6395 | |
| 3 | 0.710 | F 0.83 | 16.2% | 255402 | Acetylated F peak |
| 4 | 0.845 | F 0.99 | 64.9% | 1026017 | Consistent with F |
| 5 | 1.228 | A 1.00 | 17.4% | 275411 | A peak |
| Total Area: | | | | 1581073 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range