


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1988, Rack A, Plate 1, Well D11, 482078  
 [DC67A9B5D8EDD31D] Nov 21, 2025 12:13:05 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.47	1.2%	15386	
2	0.495	F 0.58	0.7%	8256	
3	0.713	F 0.83	15.6%	192648	Acetylated F peak
4	0.848	F 0.99	63.4%	783102	Consistent with F
5	1.233	A 1.01	19.0%	234916	A peak
Total Area:				1234308	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range