


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1380, Rack A, Plate 1, Well A06, 65073
[C9DCB74EE3FA486A] Apr 17, 2024 10:34:44 Pressure = 80 bar (80 to 80)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.392	F 0.46	2.0%	13136	
2	0.478	F 0.56	1.6%	10466	
3	0.747	F 0.88	17.1%	113248	Acetylated F peak
4	0.863	F 1.02	66.2%	438125	Consistent with F
5	1.218	A 1.01	13.2%	87131	A peak
Total Area:				662106	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range