


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2095, Rack A, Plate 1, Well F01, 498458  
 [98377F62D9E686C8] Mar 17, 2026 13:34:17 Pressure = 55 bar (55 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.422	F 0.48	0.7%	13741	
2	0.518	F 0.59	0.6%	12541	
3	0.728	F 0.83	16.9%	348372	Acetylated F peak
4	0.860	F 0.98	69.7%	1439472	Consistent with F
5	1.015	F 1.15	0.8%	15808	
6	1.258	A 1.00	11.4%	234448	A peak
Total Area:				2064382	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range