


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2191, Rack A, Plate 1, Well A10, 486281
[C92CD16AECF7BB1D] Jul 06, 2026 11:58:41 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.673	F	0.81	13.2%	129240	Acetylated F peak
2	0.820	F	0.99	59.2%	578691	Consistent with F
3	1.182	A	1.01	27.5%	269094	A peak - REVIEW
Total Area: 977025						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range