


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2169, Rack A, Plate 1, Well D04, 501693  
[C96DA8FE473A6C6A] Jun 05, 2026 14:32:36 Pressure = 65 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.508	F 0.60	1.0%	21142	
2	0.702	F 0.83	13.6%	296201	Acetylated F peak
3	0.830	F 0.98	63.0%	1367085	Consistent with F
4	1.190	A 1.01	22.4%	486672	A peak - REVIEW
Total Area:				2171100	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range