


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1359, Rack A, Plate 1, Well H11, 27403
[998CF10389279D1F] Apr 03, 2024 12:49:04

Pressure = 59 bar (59 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.47	1.5%	26492	
2	0.483	F 0.54	2.8%	48253	
3	0.765	F 0.85	8.8%	152989	Acetylated F peak
4	0.880	F 0.98	73.7%	1281502	Consistent with F
5	1.252	A 1.00	13.2%	229526	A peak
Total Area:				1738762	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range