


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1924, Rack A, Plate 1, Well C08, 170346
[9824425329D6F317] Oct 01, 2025 11:46:26 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	2.2%	13511	
2	0.488	F 0.58	1.4%	8804	
3	0.685	F 0.81	10.1%	61223	Acetylated F peak
4	0.832	F 0.99	54.4%	331069	Consistent with F
5	0.967	F 1.15	2.1%	12567	
6	1.190	A 1.00	29.8%	181124	A peak - REVIEW
Total Area:				608298	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range