


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2140, Rack A, Plate 1, Well E04, 173257
[9D221A6ACE3E8642] May 06, 2026 13:12:50 Pressure = 64 bar (63 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.425	F 0.50	0.9%	8476	
2	0.515	F 0.60	0.8%	6921	
3	0.722	F 0.84	16.7%	150136	Acetylated F peak
4	0.850	F 0.99	67.4%	604851	Consistent with F
5	1.235	A 1.01	14.1%	126565	A peak
Total Area:				896949	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range