


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2189, Rack A, Plate 1, Well H01, 173677-mmx
 [C838918275804C4A] Jul 01, 2026 15:39:44 Pressure = 81 bar (78 to 83)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|------|-------|--------|--------|--------------------|---------------------|
| 1 | 0.425 | F 0.23 | 7.5% | 37946 | |
| 2 | 0.542 | F 0.30 | 13.8% | 70172 | |
| 3 | 0.888 | F 0.49 | 33.1% | 168133 | |
| 4 | 1.760 | F 0.97 | 2.0% | 10063 | Consistent with F 6 |
| 5 | 1.862 | F 1.03 | 22.5% | 114069 | Consistent with F |
| 6 | 2.647 | A 0.76 | 2.4% | 12076 | |
| 7 | 2.970 | A 0.85 | 0.3% | 1499 | |
| 8 | 3.502 | A 1.01 | 18.5% | 93796 | A peak |
| | | | | Total Area: 507754 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range