


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2149, Rack A, Plate 1, Well C03, 501882  
[896CF552F472644A] May 20, 2026 12:59:12 Pressure = 65 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.425	F 0.50	0.9%	9835	
2	0.517	F 0.61	0.5%	5501	
3	0.715	F 0.84	11.9%	137193	Acetylated F peak
4	0.850	F 1.00	56.6%	651649	Consistent with F
5	1.223	A 1.02	30.1%	346224	A peak - REVIEW

Total Area: 1150402

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range