


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2179, Rack A, Plate 1, Well E04, 502250
[9D29FF942A3AC6CA] Jun 18, 2026 14:19:04 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.508	F	0.61	2.1%	16746	
2	0.693	F	0.83	15.9%	129476	Acetylated F peak
3	0.833	F	0.99	74.9%	610252	Consistent with F
4	1.185	A	1.00	7.2%	58307	A peak
Total Area: 814781						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range