


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2065, Rack A, Plate 1, Well F09, 172160
[9C6221BF987159B5] Feb 04, 2026 13:59:43 Pressure = 85 bar (85 to 87)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.518	F 0.62	1.1%	13804	
2	0.700	F 0.83	17.3%	215807	Acetylated F peak
3	0.847	F 1.01	71.9%	899085	Consistent with F
4	1.187	A 1.00	9.8%	122164	A peak
Total Area:				1250860	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range