


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2080, Rack A, Plate 1, Well F03, 489576
[8C727352F60EAC60] Feb 24, 2026 12:33:11 Pressure = 56 bar (56 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.47	1.4%	14481	
2	0.513	F 0.58	0.7%	7130	
3	0.732	F 0.83	16.4%	164025	Acetylated F peak
4	0.868	F 0.98	67.8%	678355	Consistent with F
5	1.018	F 1.16	1.3%	13053	
6	1.263	A 1.01	12.4%	124137	A peak
Total Area: 1001181					

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range