


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1982, Rack A, Plate 1, Well D02, 482045
[D966A252F42DFBB7] Nov 19, 2025 11:52:30 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.47	0.9%	13158	
2	0.502	F 0.58	0.7%	9670	
3	0.723	F 0.84	18.4%	257202	Acetylated F peak
4	0.847	F 0.98	71.4%	998044	Consistent with F
5	1.012	F 1.17	0.8%	11626	
6	1.242	A 1.01	7.8%	108767	A peak
Total Area:				1398467	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range