


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2185, Rack A, Plate 1, Well C03, 486025
[D938AA567630C4E8] Jun 25, 2026 12:39:09 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.428	F 0.51	0.5%	4662	
2	0.510	F 0.61	0.3%	2982	
3	0.688	F 0.82	13.2%	114714	Acetylated F peak
4	0.832	F 0.99	74.9%	649574	Consistent with F
5	1.185	A 1.01	11.0%	95868	A peak
Total Area:				867800	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range