


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1813, Rack A, Plate 1, Well A03, 457135 FAS
 [8820A85B9CCF2C60] May 22, 2025 16:34:15 Pressure = 60 bar (58 to 61)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.21	2.2%	132527	
2	0.543	F 0.29	3.6%	213417	
3	0.940	F 0.50	29.0%	1732320	
4	1.847	F 0.97	49.6%	2962452	Consistent with F
5	2.442	F 1.29	0.7%	39439	
6	2.672	F 1.41	0.8%	46027	
7	2.947	A 0.79	0.5%	27712	
8	3.213	A 0.86	0.3%	15780	
9	3.732	A 1.00	7.2%	432731	A peak
10	4.207	A 1.12	0.4%	22979	
11	5.327	S 1.01	5.8%	346163	Consistent with S

Total Area: 5971547

A/V=1.24

* * * The total peak area is too low/high for evaluation! * * *

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range