


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1818, Rack A, Plate 1, Well B05, 455888  
[8D70EECB176F2CEA] May 28, 2025 09:57:32 Pressure = 61 bar (61 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.397	F 0.47	1.6%	13377	
2	0.492	F 0.58	1.1%	9007	
3	0.733	F 0.86	12.7%	106078	Acetylated F peak
4	0.848	F 1.00	56.4%	471653	Consistent with F
5	0.988	F 1.17	2.0%	16401	
6	1.228	A 1.01	26.2%	219024	A peak - REVIEW
Total Area:				835540	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range