


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2025, Rack A, Plate 1, Well A04, 482248
[9C76847DC4F7D9B5] Dec 23, 2025 11:09:03 Pressure = 78 bar (77 to 80)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	0.8%	5023	
2	0.500	F 0.59	0.7%	4777	
3	0.707	F 0.83	13.4%	88766	Acetylated F peak
4	0.843	F 0.99	62.2%	412351	Consistent with F
5	1.000	F 1.18	2.3%	14982	
6	1.227	A 1.02	20.7%	137237	A peak - REVIEW
Total Area:				663136	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range