


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1801, Rack A, Plate 1, Well C03, 455887
 [D874B947A13AD19D] May 21, 2025 10:43:53 Pressure = 60 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.48	1.0%	9325	
2	0.492	F 0.58	1.1%	10166	
3	0.703	F 0.83	12.9%	124569	Acetylated F peak
4	0.853	F 1.01	55.4%	536030	Consistent with F
5	0.982	F 1.16	2.8%	26879	
6	1.093	A 0.90	0.2%	2155	
7	1.232	A 1.02	26.7%	257852	A peak - REVIEW
Total Area:				966976	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range