


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1998, Rack A, Plate 1, Well A04, 171197 FA  
 [D976FD67AE6DD19F] Nov 28, 2025 10:54:32 Pressure = 72 bar (72 to 74)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.395	F 0.46	1.4%	10151	
2	0.497	F 0.58	0.7%	4681	
3	0.700	F 0.82	14.1%	100680	Acetylated F peak
4	0.853	F 0.99	64.5%	460856	Consistent with F
5	1.007	F 1.17	1.7%	11885	
6	1.242	A 1.01	17.7%	126510	A peak
Total Area:				714763	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range