


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2077, Rack A, Plate 1, Well F08, 504548  
[D82326D2F5597117] Feb 18, 2026 12:05:54 Pressure = 55 bar (54 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.47	1.5%	14848	
2	0.512	F 0.58	0.6%	6070	
3	0.727	F 0.82	13.7%	137920	Acetylated F peak
4	0.870	F 0.98	63.4%	637477	Consistent with F
5	1.257	A 1.00	20.7%	208437	A peak - REVIEW
Total Area:				1004752	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range