


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2038, Rack A, Plate 1, Well C06, 489997  
 [DD3398A307B7FB9D] Jan 09, 2026 11:37:29 Pressure = 80 bar (80 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	1.3%	10615	
2	0.497	F 0.59	0.9%	7217	
3	0.702	F 0.83	9.0%	70737	Acetylated F peak
4	0.842	F 1.00	64.4%	506589	Consistent with F
5	1.002	F 1.19	1.2%	9759	
6	1.213	A 1.01	23.1%	182216	A peak - REVIEW
Total Area:				787133	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range