


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2204, Rack A, Plate 1, Well C02, 496030  
[992CC9F94597B317] Jul 23, 2026 11:01:38 Pressure = 93 bar (93 to 95)

FA



| PEAK        | RT    | REL | RT   | % CONC | AREA   | COMMENT           |
|-------------|-------|-----|------|--------|--------|-------------------|
| 1           | 0.515 | F   | 0.62 | 1.2%   | 10392  |                   |
| 2           | 0.688 | F   | 0.83 | 11.4%  | 99405  | Acetylated F peak |
| 3           | 0.825 | F   | 1.00 | 67.7%  | 592148 | Consistent with F |
| 4           | 1.157 | A   | 1.00 | 19.8%  | 173287 | A peak            |
| Total Area: |       |     |      |        | 875232 |                   |

Codes: 1) Wide A peak  
2) Area of A peak < 80%  
3) Peak area greater than expected  
4) Peak after A2  
5) A1c > 10%  
6) HbF or variant present  
7) Total sample area too small/big  
8) A2 is not within normal range