


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2188, Rack A, Plate 1, Well B04, 493027  
[D83981D300F7B997] Jul 01, 2026 11:22:50 Pressure = 79 bar (79 to 82)

FA



| PEAK        | RT    | REL RT | % CONC | AREA    | COMMENT           |
|-------------|-------|--------|--------|---------|-------------------|
| 1           | 0.683 | F 0.82 | 20.1%  | 369286  | Acetylated F peak |
| 2           | 0.815 | F 0.97 | 70.8%  | 1300584 | Consistent with F |
| 3           | 1.170 | A 1.01 | 9.1%   | 167629  | A peak            |
| Total Area: |       |        |        | 1837499 |                   |

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range