


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well B04, 493096
[983992CAEA20C64A] Jul 15, 2026 12:47:12 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.515	F	0.62	0.2%	2951	
2	0.683	F	0.82	11.8%	166778	Acetylated F peak
3	0.817	F	0.99	59.0%	831876	Consistent with F
4	1.158	A	1.00	29.0%	408531	A peak - REVIEW
Total Area: 1410136						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range