


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2200, Rack A, Plate 1, Well E02, 173986
[DD6C9238997519BD] Jul 17, 2026 15:57:01 Pressure = 82 bar (82 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.423	F	0.50	2.1%	10007	
2	0.520	F	0.62	0.2%	869	
3	0.678	F	0.81	16.1%	75389	Acetylated F peak
4	0.832	F	0.99	63.4%	297002	Consistent with F
5	1.177	A	1.02	18.2%	85003	A peak
Total Area:					468270	

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range