


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2080, Rack A, Plate 1, Well F02, 489575
[98732607A35B79B7] Feb 24, 2026 12:31:10 Pressure = 56 bar (56 to 58)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.417	F	0.47	0.9%	11121	
2	0.517	F	0.59	0.7%	8597	
3	0.732	F	0.83	14.9%	184840	Acetylated F peak
4	0.868	F	0.98	66.6%	824820	Consistent with F
5	1.262	A	1.01	16.9%	208901	A peak
Total Area: 1238279						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range