


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1982, Rack A, Plate 1, Well E07, 482048  
 [9C26A307A1F8ACEA] Nov 19, 2025 12:26:52 Pressure = 68 bar (67 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.47	0.5%	9615	
2	0.497	F 0.58	0.5%	9007	
3	0.718	F 0.83	12.3%	245391	Acetylated F peak
4	0.855	F 0.99	69.3%	1377770	Consistent with F
5	1.245	A 1.01	17.5%	347404	A peak
Total Area:				1989187	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range