


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2173, Rack A, Plate 1, Well C06, 173659
[8C29A827846F9117] Jun 10, 2026 14:02:10 Pressure = 66 bar (66 to 68)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.49	2.8%	9788	
2	0.500	F 0.59	2.5%	8622	
3	0.690	F 0.82	7.3%	25462	Acetylated F peak
4	0.855	F 1.01	28.5%	99737	Consistent with F
5	0.983	F 1.17	4.7%	16468	
6	1.195	A 0.99	54.2%	189690	A peak - REVIEW

Total Area: 349767

* * * The total peak area is too low/high for evaluation! * * *

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range