


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1787, Rack A, Plate 1, Well B05, 457117
[DC60B2EC32A584C2] Apr 30, 2025 13:45:14 Pressure = 90 bar (88 to 91)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.50	2.2%	13561	
2	0.490	F 0.61	1.2%	7476	
3	0.708	F 0.88	21.9%	133693	
4	0.830	F 1.03	69.3%	422984	Consistent with F
5	1.160	A 1.01	5.3%	32246	A peak
Total Area:				609960	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range