


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2127, Rack A, Plate 1, Well A06, 498271
[8D234D9F9FB484C8] Apr 21, 2026 09:54:49 Pressure = 65 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.520	F 0.60	1.9%	22274	
2	0.715	F 0.83	17.3%	198033	Acetylated F peak
3	0.863	F 1.00	68.2%	779242	Consistent with F
4	1.233	A 1.00	12.6%	143711	A peak
Total Area:				1143260	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range