


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2022, Rack A, Plate 1, Well F07, 171570 FA  
 [983681C2DF820EC8] Dec 19, 2025 12:23:21 Pressure = 74 bar (74 to 76)



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	0.9%	12087	
2	0.492	F 0.57	0.5%	7427	
3	0.745	F 0.87	9.5%	132399	
4	0.843	F 0.98	57.6%	802791	Acetylated F peak Consistent with F
5	1.013	F 1.18	0.6%	8994	
6	1.232	A 1.01	30.8%	429841	A peak - REVIEW
Total Area:				1393539	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range