


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2038, Rack A, Plate 1, Well C03, 489994
[88738DF652EAA66A] Jan 09, 2026 11:29:24 Pressure = 81 bar (80 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.49	0.9%	9505	
2	0.500	F 0.60	0.7%	7248	
3	0.705	F 0.84	17.5%	191666	Acetylated F peak
4	0.857	F 1.02	65.1%	713588	Consistent with F
5	1.208	A 1.02	15.9%	174328	A peak
Total Area:				1096335	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range