


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2188, Rack A, Plate 1, Well A12, 493023
[992CC1D300FFB13D] Jul 01, 2026 11:14:45 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.512	F 0.61	1.6%	16669	
2	0.707	F 0.84	9.9%	104631	Acetylated F peak
3	0.823	F 0.98	78.8%	835329	Consistent with F
4	1.160	A 1.00	9.8%	103641	A peak

Total Area: 1060270

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range