


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1884, Rack A, Plate 1, Well B04, 457245
[8C20324081D984CA] Aug 13, 2025 11:13:44 Pressure = 61 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.49	0.7%	10090	
2	0.498	F 0.59	1.0%	13454	
3	0.705	F 0.84	14.6%	203439	Acetylated F peak
4	0.830	F 0.99	69.4%	965405	Consistent with F
5	1.217	A 1.01	14.3%	199122	A peak
Total Area:				1391510	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range