


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2095, Rack A, Plate 1, Well E04, 498817
 [89333A378C99F3BD] Mar 17, 2026 13:14:05 Pressure = 56 bar (55 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.47	1.3%	15130	
2	0.517	F 0.59	0.7%	7942	
3	0.727	F 0.83	16.7%	187909	Acetylated F peak
4	0.873	F 0.99	69.5%	781962	Consistent with F
5	1.253	A 1.00	11.8%	132247	A peak
Total Area:				1125190	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range