


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2025, Rack A, Plate 1, Well B10, 171602  
[DD27856DEE755337] Dec 23, 2025 11:45:26 Pressure = 77 bar (76 to 78)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.390	F 0.46	2.2%	13796	
2	0.488	F 0.58	1.5%	9448	
3	0.698	F 0.82	5.7%	35971	Acetylated F peak
4	0.848	F 1.00	34.1%	216057	Consistent with F
5	1.010	F 1.19	3.6%	22940	
6	1.225	A 1.01	52.1%	330289	A peak - REVIEW
7	1.395	S 0.89	0.9%	5962	
Total Area:				634463	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range