


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2034, Rack A, Plate 1, Well H06, 490365
 [C82692467C208CC2] Jan 06, 2026 13:42:34 Pressure = 82 bar (82 to 84)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.410	F 0.48	0.7%	11032	
2	0.502	F 0.59	0.4%	6339	
3	0.703	F 0.82	15.3%	249568	Acetylated F peak
4	0.832	F 0.97	72.8%	1186544	Consistent with F
5	1.198	A 1.01	10.8%	175762	A peak
Total Area:				1629245	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range